

MINIMED™ INFUSION SETS

A SET FOR EVERY BODY
A SET FOR EVERY LIFESTYLE

Learn more about MiniMed™ Mio™ Advance at www.medtronic-diabetes.co.uk

	Easy	Compact		30° Insertion Angle	Comfortable		Adjustable Insertion Angle		Secure	
	MiniMed™ Mio™ Advance	MiniMed™ Mio™		MiniMed™ Mio™ 30	MiniMed™ Quick-Set™		MiniMed™ Silhouette™		MiniMed™ Sure-T™	
	Convenience with an Integrated Serter									
										
User Needs & Special Circumstances ^{1,2}	<ul style="list-style-type: none"> ▪ Dexterity issues ▪ Needle anxiety 	<ul style="list-style-type: none"> ▪ On-the-go lifestyle ▪ Choice of three colours 		<ul style="list-style-type: none"> ▪ Lean/ Active ▪ History of bent cannula 	<ul style="list-style-type: none"> ▪ Universal fit 		<ul style="list-style-type: none"> ▪ Lean/active ▪ History of bent cannula 		<ul style="list-style-type: none"> ▪ History of bent cannula ▪ History of scar tissue ▪ Teflon allergy 	
Body Type	Medium/ high BMI*	Medium/ high BMI		Low BMI	Medium/ high BMI		Low BMI		Medium/ high BMI	
Insertion Device	(Integrated) serter	(Integrated) serter		(Integrated) serter	MiniMed Quick-serter		MiniMed Sil-serter		Manual	
Cannula/ Needle Type	Soft	Soft		Soft	Soft		Soft		Steel	
Insertion Angle	90°	90°		30°	90°		20°-45°		90°	
Cannula Length	6 mm	6 mm	9 mm	13 mm	6 mm	9 mm	13 mm	17 mm	6 mm, 8 mm, 10 mm	
Cannula Fill	0.6 Units	0.3 Units	0.5 Units	0.7 Units	0.3 Units	0.5 Units	0.7 Units	0.7 Units	No cannula fill required	
Tubing Length	60 cm	45 cm, 60 cm, 80 cm		60 cm, 110 cm	45 cm, 60 cm, 80 cm, 110 cm		45 cm, 60 cm, 80 cm, 110 cm		45 cm, 60 cm, 80 cm	
Cannula Thickness (Gauge)	25G	25G		25G	25G		25G		29G	
Change Frequency	2-3 days	2-3 days		2-3 days	2-3 days		2-3 days		1-2 days	

1. Medtronic Infusion Set Advisory Board 2016, Northridge, CA. Final report on file at Medtronic.

2. Helpline Complaint Data Report 2014, 2015, 2016. Medtronic data on file.

Important Safety Information: Important Safety Information: Infusion sets are indicated for the subcutaneous infusion of medication, including insulin, from infusion pumps. They are neither intended nor indicated for intravenous (IV) infusion or the infusion of blood or blood products. Inaccurate medication, delivery, infection and/or site irritation may result from improper insertion and maintenance of the infusion insertion and maintenance of the infusion site. Replace the infusion set every 48-72 hours, or per your healthcare professional's instructions. For a listing of indications, contraindications, precautions, warnings, and potential adverse events, please refer to the Instructions for Use.

Legal manufacturer: Unomedical a/s, Aaholmvej 1-3, Osted. 4320 Lejre, Denmark

UC202000376EN - May 2019

Medtronic
Further, Together